



De-Mountable 1" OD Pole Assembly with Top Flange to Customer Requirements & G1/2" or G3/4" Male Thread for Sensor Connection

Max Length for PSM Supplied Pole Assembly 2000mm. Alternatively Yard to Fabricate

Assembled Length From Underside of Top Flange to End of Sensor TBC

APT Transmitter with G1/2" or G3/4" Female Pole Adaptor
Drawing: S-001477-APT-DD

Sensor Support Bracket / Pipe
(SY Supply)

Recommended 50mm min Clearance to Tank Bottom

Material Options:
All 316 ST/STL
Hastelloy / Tantalum Process Connection / Stainless Body

For Cable Termination Refer to Drawing: S-001479-APT-GA



INSTRUMENTATION LTD
PART OF THE **scanjet** GROUP

TITLE **APT Submersible Pole Assembly Installation**
Datasheets
APT500 DAT054 / APT1000 DAT57

DRAWING No
S-001493-APT-DD
SALES DRAWING ONLY

SHT 1 OF 1
A4

SCALE	NTS	MATERIAL AND SPECIFICATION
	FINISH	
ALL DIMENSIONS ARE IN MILLIMETERS UNLESS OTHERWISE STATED TOLERANCES UNLESS OTHERWISE STATED +/- 0.1 ANG #/2	B	CHECKED
	DATE	
	TR	
	DRAWN	
ISSUE	18-11-19	SAL
		MU